REVARON glovebox maintenance log sheet

Data and time to be checked	Status (Working/ Not Working)	Box Pressure (mBar)	H₂O (ppm)	O ₂ (ppm)	Freezer Temp (°C)	Ar Cylinder Pressure (Bar)	Ar:H ₂ Cylinder Pressure (Bar)	Rotary Pump Oil Level	Rotary Pump Oil Colour	Freezer (Working/ Not Working)	Regeneration (done / not done)	Comments
12/4/2099	Working	3mBar	0.8ppm	0.5ppm	-3 °C	5.5 Bar	5 Bar	70%	Yellow	Working	Not done	Entered by XYZ

For sales, support and maintenance please contact info@revaron.com, Mobile and WhatsApp: +91 7676564755

